



# Type examination certificate

## No. CH-MI001-09027-05

*Bauartprüfzertifikat*

*Applicant:*  
*Auftraggeber:*

**G2 MISURATORI SRL**  
Via san Martino 38  
14100 Asti  
Italy

*Requirements:*  
*Anforderungen:*

Swiss ordinance on measuring instruments (SR  
941.210) of February 15, 2006 annex 2 module B;

Directive 2004/22/EC of the European Parliament and  
Council of March 31, 2004 on measuring instruments  
(MID), appendix B, class of measuring instruments  
MI-001.

Conformity of superseded standard:  
EN 14154 Edition 2007

*Type of instrument:*  
*Geräteart:*

**Multi-jet impeller meter intended for the metering of  
cold water**

*Type designation:*  
*Typenbezeichnung:*

**AFM, AFM-P, AFML, AFM-V  
BFM, BFM-P, BFMLH, BFML, BFMLR2, BFM-V  
PFM, PFM-P, PFMLH, PFML, PFMLR2, PFM-V  
QDFM, QDFM-V, QDFM-P, QDFML, QDFMLH, QDFMLR2**

*Accuracy class(es):*  
*Genauigkeitsklasse(n):*

2 (according to OIML R 49:2006)

*Characteristics:*  
*Kenndaten:*

PN 16, DN 15 ... DN 50, T30, T50, Class B, E1  
Q<sub>3</sub> 1.6 ... Q<sub>3</sub> 25.0

*Certificate valid until:*  
*Zertifikat gültig bis:*

July 22, 2019

CH-3003 Bern-Wabern, November 02, 2010

*Notified body:*  
*Benannte Stelle:*

Certification Body METAS-Cert  
No. 1259

*For the test:*  
*Für die Prüfung:*

Dr. Hugo Bissig

Jürg Ramseyer, Head of METAS-Cert